



# Assay Calibration Report

## Assay Information

Assay Name: Cylindrospermopsin 1X Units: ng/mL  
 Assay Mode: 4-Parameter Logistic # of decimals: 3  
 Normal: 0.050 - 2.000 Assay Description:

## Controls:

Normal Control


## Standards:

Std1, Concentration = 0.000, Minimum number to use: 2  
 Std2, Concentration = 0.050, Minimum number to use: 2  
 Std3, Concentration = 0.100, Minimum number to use: 2  
 Std4, Concentration = 0.250, Minimum number to use: 2  
 Std5, Concentration = 0.500, Minimum number to use: 2  
 Std6, Concentration = 1.000, Minimum number to use: 2  
 Std7, Concentration = 2.000, Minimum number to use: 2  
 Curve valid interval: 7 days 0 hours  
 Axis Mode: Y = Abs, X = Log(Conc)

## Assay Calibration and Statistics

Name	Absorbance	Concentration	Position
7/10/2012 11:53:44 AM			
Std1	1.476 Abs	0.004 ng/mL	A01
Std1	1.502 Abs	< 0.000 ng/mL	B01
Std2	1.356 Abs	0.042 ng/mL	C01
Std2	1.328 Abs	0.051 ng/mL	D01
Std3	1.168 Abs	0.113 ng/mL	E01
Std3	1.168 Abs	0.113 ng/mL	F01
Std4	0.852 Abs	0.319 ng/mL	G01
Std4	1.083 Abs	0.154 ng/mL	H01
Std5	0.689 Abs	0.535 ng/mL	A02
Std5	0.684 Abs	0.544 ng/mL	B02
Std6	0.531 Abs	0.986 ng/mL	C02
Std6	0.523 Abs	1.024 ng/mL	D02
Std7	0.413 Abs	1.942 ng/mL	E02
Std7	0.414 Abs	1.928 ng/mL	F02
7/10/2012 11:53:44 AM			
Normal Control	0.605 Abs	0.724 ng/mL	G02
Normal Control	0.618 Abs	0.689 ng/mL	H02

Name	Mean Abs	SD Abs	CV Abs	Mean Conc	SD Conc	CV Conc	Diff Conc
Std1	1.489	0.018	1.23				
Std2	1.342	0.020	1.48	0.046	0.006	13.69	-8.00
Std3	1.168	0.000	0.00	0.113	0.000	0.00	13.00
Std4	0.967	0.163	16.88	0.236	0.117	49.33	-5.60
Std5	0.687	0.004	0.52	0.539	0.006	1.18	7.80
Std6	0.527	0.006	1.07	1.005	0.027	2.67	0.50
Std7	0.414	0.001	0.17	1.935	0.010	0.51	-3.25





## Assay Calibration Report

### Assay Information

Assay Name: Cylindrospermopsin 1X Units: ng/mL  
Assay Mode: 4-Parameter Logistic # of decimals: 3  
Normal: 0.050 - 2.000 Assay Description:


Controls:  
Normal Control

#### Standards:

Std1, Concentration = 0.000, Minimum number to use: 2  
Std2, Concentration = 0.050, Minimum number to use: 2  
Std3, Concentration = 0.100, Minimum number to use: 2  
Std4, Concentration = 0.250, Minimum number to use: 2  
Std5, Concentration = 0.500, Minimum number to use: 2  
Std6, Concentration = 1.000, Minimum number to use: 2  
Std7, Concentration = 2.000, Minimum number to use: 2  
Curve valid interval: 7 days 0 hours  
Axis Mode: Y = Abs, X = Log(Conc)

### Assay Calibration and Statistics

Name	Mean Abs	SD Abs	CV Abs	Mean Conc	SD Conc	CV Conc	Diff Conc
Normal Control	0.612	0.009	1.50	0.706	0.025	3.50	





# Test Report

## Test Information

Name/ID	Assay	Absorbance	Concentration	Interpretation	Reference	Position
7/10/2012 11:53:44 AM						
Std1	Cylindrospermopsin 1X	1.476 Abs	0.004 ng/mL			A01
Std1	Cylindrospermopsin 1X	1.502 Abs	< 0.000 ng/mL			B01
Std2	Cylindrospermopsin 1X	1.356 Abs	0.042 ng/mL			C01
Std2	Cylindrospermopsin 1X	1.328 Abs	0.051 ng/mL			D01
Std3	Cylindrospermopsin 1X	1.168 Abs	0.113 ng/mL			E01
Std3	Cylindrospermopsin 1X	1.168 Abs	0.113 ng/mL			F01
Std4	Cylindrospermopsin 1X	0.852 Abs	0.319 ng/mL			G01
Std4	Cylindrospermopsin 1X	1.083 Abs	0.154 ng/mL			H01
Std5	Cylindrospermopsin 1X	0.689 Abs	0.535 ng/mL			A02
Std5	Cylindrospermopsin 1X	0.684 Abs	0.544 ng/mL			B02
Std6	Cylindrospermopsin 1X	0.531 Abs	0.986 ng/mL			C02
Std6	Cylindrospermopsin 1X	0.523 Abs	1.024 ng/mL			D02
Std7	Cylindrospermopsin 1X	0.413 Abs	1.942 ng/mL			E02
Std7	Cylindrospermopsin 1X	0.414 Abs	1.928 ng/mL			F02
Normal Control	Cylindrospermopsin 1X	0.605 Abs	0.724 ng/mL			G02
Normal Control	Cylindrospermopsin 1X	0.618 Abs	0.689 ng/mL			H02
AB09471	Cylindrospermopsin 1X	1.620 Abs	< 0.000 ng/mL	Out(LR)	0.050 - 2.000	A03
AB09471	Cylindrospermopsin 1X	1.579 Abs [1.5995]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	B03
AB09472	Cylindrospermopsin 1X	1.636 Abs	< 0.000 ng/mL	Out(LR)	0.050 - 2.000	C03
AB09472	Cylindrospermopsin 1X	1.551 Abs [1.5935]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	D03
AB09469	Cylindrospermopsin 1X	1.601 Abs	< 0.000 ng/mL	Out(LR)	0.050 - 2.000	E03
AB09469	Cylindrospermopsin 1X	1.570 Abs [1.5855]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	F03
AB09470	Cylindrospermopsin 1X	3.785 Abs	< 0.000 ng/mL	Out(A,LR)	0.050 - 2.000	G03
AB09470	Cylindrospermopsin 1X	1.586 Abs [2.6855]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	H03
AB09473	Cylindrospermopsin 1X	1.532 Abs	< 0.000 ng/mL	Out(LR)	0.050 - 2.000	A04
AB09473	Cylindrospermopsin 1X	1.634 Abs [1.5830]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	B04
20120709LB	Cylindrospermopsin 1X	1.531 Abs	< 0.000 ng/mL	Out(LR)	0.050 - 2.000	C04
20120709LB	Cylindrospermopsin 1X	1.582 Abs [1.5565]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	D04
AB09472LD	Cylindrospermopsin 1X	1.554 Abs	< 0.000 ng/mL	Out(LR)	0.050 - 2.000	E04
AB09472LD	Cylindrospermopsin 1X	1.574 Abs [1.5640]	< 0.000 ng/mL [< 0.000]	Out(LR) [Out(LR)]	0.050 - 2.000	F04

Notes

Signature

*Reddy*  
7/10/2012



## Cylindrospermopsin 1X ELISA Summary Report

Office of Water Quality - Watershed Assessment and Planning Branch

Sample #	Location	Date Collected	Date Analyzed	Conc. (ppb)
AB09471	Salamonie Reservoir-Lost Bridge W SRA	7/9/2012	7/10/2012	<0.050
AB09472	Worster Lake-Potato Creek State Park	7/9/2012	7/10/2012	<0.050
AB09469	Field Duplicate (Salamonie)	7/9/2012	7/10/2012	<0.050
AB09470	Field Blank	7/9/2012	7/10/2012	<0.050
AB09473	Sand Lake-Chain O'Lakes State Park	7/9/2012	7/10/2012	<0.050
20120709LB	Lab Blank	7/9/2012	7/10/2012	<0.050
AB09472LD	Lab Duplicate (Worster Lake)	7/9/2012	7/10/2012	<0.050