



September 11, 2019

Arcelor Mittal USA, Inc.
250 W US Highway 12
Burns Harbor, IN 46304-9745

Work Order No.: 19H1769

Re: Daily

Dear Teri Kirk:

Microbac Laboratories, Inc. - Chicagoland Division received 19 sample(s) on 8/28/2019 10:15:00AM for the analyses presented in the following report as Work Order 19H1769.

The enclosed results were obtained from and are applicable to the sample(s) as received at the laboratory. All sample results are reported on an "as received" basis unless otherwise noted.

All data included in this report have been reviewed and meet the applicable project specific and certification specific requirements, unless otherwise noted. A qualifications page is included in this report and lists the programs under which Microbac maintains certification.

This report has been paginated in its entirety and shall not be reproduced except in full, without the written approval of Microbac Laboratories.

We appreciate the opportunity to service your analytical needs. If you have any questions, please contact your project manager. For any feedback, please contact Ron Misiunas, Division Manager, at ron.misiunas@microbac.com.

Sincerely,
Microbac Laboratories, Inc.

A handwritten signature in black ink that reads "Carey Gadzala". The signature is written in a cursive, flowing style.

Carey Gadzala
Project Manager

[Microbac Laboratories, Inc.](#)



WORK ORDER SAMPLE SUMMARY

Date: *Wednesday, September 11, 2019*

Client: Arcelor Mittal USA, Inc.
Project: Daily
Lab Order: 19H1769

Lab Sample ID	Client Sample ID	Tag Number	Collection Date	Date Received
19H1769-01	011-Composite	011	08/27/2019 04:59	8/28/2019 10:15:00AM
19H1769-02	011-Grab	011	08/27/2019 04:59	8/28/2019 10:15:00AM
19H1769-03	001-Composite	001	08/27/2019 05:20	8/28/2019 10:15:00AM
19H1769-04	001-Grab	001	08/27/2019 05:20	8/28/2019 10:15:00AM
19H1769-05	031-Grab	031	08/28/2019 05:37	8/28/2019 10:15:00AM
19H1769-06	Mixed Liquor-Grab	Mixed Liquor	08/28/2019 05:40	8/28/2019 10:15:00AM
19H1769-07	J-Box-Grab	J-Box	08/28/2019 05:35	8/28/2019 10:15:00AM
19H1769-08	WWII-Grab	WWII	08/28/2019 06:10	8/28/2019 10:15:00AM
19H1769-09	Coldwell-Grab	Coldwell	08/28/2019 06:20	8/28/2019 10:15:00AM
19H1769-10	RSB FT Overflow-Grab	RSB FT Overflow	08/28/2019 06:25	8/28/2019 10:15:00AM
19H1769-11	RSB FT Influent-Grab	RSB FT Influent	08/28/2019 06:26	8/28/2019 10:15:00AM
19H1769-12	BFTD-Grab	BFTD	08/28/2019 06:48	8/28/2019 10:15:00AM
19H1769-13	WPL-Grab	WPL	08/26/2019 07:58	8/28/2019 10:15:00AM
19H1769-14	999-Grab	999	08/28/2019 06:32	8/28/2019 10:15:00AM
19H1769-15	BFTC-Grab	BFTC	08/28/2019 06:52	8/28/2019 10:15:00AM
19H1769-16	002-Composite	002	08/27/2019 06:56	8/28/2019 10:15:00AM
19H1769-17	002-Grab	002	08/27/2019 06:56	8/28/2019 10:15:00AM
19H1769-18	WAL-Grab	WAL	08/27/2019 07:10	8/28/2019 10:15:00AM

Microbac Laboratories, Inc.

250 West 84th Drive | Merrillville, IN 46410 | 800.536.8379 p | 219.769.8378 p | 219.769.1664 f | www.microbac.com

Field Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order:	19H1769
Client Project:	Daily		
Client Sample ID:	011-Grab	Work Order/ID:	19H1769-02
Sample Description:	011	Sampled:	08/27/2019 04:59
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
FLD_CL_TITR	0.00	mg/L
pH	7.7	pH Units

Client Sample ID:	001-Grab	Work Order/ID:	19H1769-04
Sample Description:	001	Sampled:	08/27/2019 05:20
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
FLD_CL_TITR	0.00	mg/L
pH	7.8	pH Units

Client Sample ID:	J-Box-Grab	Work Order/ID:	19H1769-07
Sample Description:	J-Box	Sampled:	08/28/2019 05:35
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
pH	8.6	pH Units

Client Sample ID:	RSB FT Overflow-Grab	Work Order/ID:	19H1769-10
Sample Description:	RSB FT Overflow	Sampled:	08/28/2019 06:25
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
pH	9.0	pH Units

Client Sample ID:	999-Grab	Work Order/ID:	19H1769-14
Sample Description:	999	Sampled:	08/28/2019 06:32
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
pH	8.0	pH Units

Client Sample ID:	002-Grab	Work Order/ID:	19H1769-17
Sample Description:	002	Sampled:	08/27/2019 06:56
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
pH	8.2	pH Units

Client Sample ID:	WAL-Grab	Work Order/ID:	19H1769-18
Sample Description:	WAL	Sampled:	08/27/2019 07:10
Matrix:	Aqueous	Received:	08/28/2019 10:15

Analyses	Result	Units
pH	8.9	pH Units

Field Results

Date: *Wednesday, September 11, 2019*

CASE NARRATIVE**Date:** *Wednesday, September 11, 2019***Client:** Arcelor Mittal USA, Inc.**Project:** Daily**Lab Order:** 19H1769

Report has been revised at the clients request to include Cu and Ag for Outfall 001. 9/11/19

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-01
Client Project:	Daily	Sampled:	08/27/2019 4:59
Client Sample ID:	011-Composite	Received:	08/28/2019 10:15
Sample Description:	011		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: EPA 200.7 Rev 4.4			Analyst: RPL			
Total Recoverable Metals by ICP						Prep Date/Time: 08/28/2019 11:04			
Lead	ejj	A	ND	0.0033	0.0075	U	mg/L	1	08/28/2019 13:38
Zinc	ejj	A	ND	0.0073	0.020	U	mg/L	1	08/28/2019 13:38
			Method: SM 4500-CN C/E-1999			Analyst: ABG			
Total Cyanide						Prep Date/Time: 08/28/2019 11:45			
Cyanide, Total	ejj	A	ND	0.0020	0.0050	U	mg/L	1	08/28/2019 14:38
			Method: SW-846 9014			Analyst: ABG			
Free Cyanide						Prep Date/Time: 08/28/2019 11:23			
Free Cyanide		A	ND		0.0062		mg/L	1	08/28/2019 12:30
			Method: EPA 350.1 Rev 2.0			Analyst: ABG			
Nitrogen, Ammonia as N						Prep Date/Time: 08/28/2019 11:20			
Nitrogen, Ammonia (As N)	ei	A	0.085	0.054	0.10		mg/L	1	08/28/2019 12:52
			Method: EPA 420.4 Rev 1.0			Analyst: ABG			
Total Phenolics						Prep Date/Time: 08/28/2019 11:10			
Phenolics, Total Recoverable	ejj	A	ND	0.0060	0.010	U	mg/L	1	08/28/2019 13:47
			Method: SM 2540 D-1997			Analyst: KMT			
Total Suspended Solids						Prep Date/Time: 08/28/2019 11:15			
Total Suspended Solids	ejj	A	2.0	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-02
Client Project:	Daily	Sampled:	08/27/2019 4:59
Client Sample ID:	011-Grab	Received:	08/28/2019 10:15
Sample Description:	011		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed	
			Method: EPA 1664B				Analyst: KMT			
Oil & Grease (HEM) by SPE										
Prep Date/Time: 08/28/2019 07:32										
Oil & Grease (HEM)	ejj	A	ND	1.4	5.0	U	mg/L	1	08/28/2019 14:12	

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-03
Client Project:	Daily	Sampled:	08/27/2019 5:20
Client Sample ID:	001-Composite	Received:	08/28/2019 10:15
Sample Description:	001		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
Method: EPA 200.7 Rev 4.4									
Analyst: RPL									
Prep Date/Time: 08/28/2019 11:04									
Total Recoverable Metals by ICP									
Copper	ejj	A	0.0027	0.0013	0.010		mg/L	1	08/28/2019 13:43
Lead	ejj	A	ND	0.0033	0.0075	U	mg/L	1	08/28/2019 13:43
Zinc	ejj	A	ND	0.0073	0.020	U	mg/L	1	08/28/2019 13:43
Method: EPA 200.8 Rev 5.4									
Analyst: BTM									
Prep Date/Time: 09/08/2019 12:49									
Total Recoverable Metals by ICP/MS									
Silver	ejj	A	ND	0.000053	0.00060	U	mg/L	1	09/09/2019 12:50
Method: SM 4500-CN C/E-1999									
Analyst: ABG									
Prep Date/Time: 08/28/2019 11:45									
Total Cyanide									
Cyanide, Total	ejj	A	0.0031	0.0020	0.0050		mg/L	1	08/28/2019 14:40
Method: SW-846 9014									
Analyst: ABG									
Prep Date/Time: 08/28/2019 11:23									
Free Cyanide									
Free Cyanide		A	ND		0.0062		mg/L	1	08/28/2019 12:32
Method: EPA 350.1 Rev 2.0									
Analyst: ABG									
Prep Date/Time: 08/28/2019 11:20									
Nitrogen, Ammonia as N									
Nitrogen, Ammonia (As N)	ei	A	0.27	0.054	0.10		mg/L	1	08/28/2019 12:59
Method: EPA 420.4 Rev 1.0									
Analyst: ABG									
Prep Date/Time: 08/28/2019 11:10									
Total Phenolics									
Phenolics, Total Recoverable	ejj	A	ND	0.0060	0.010	U	mg/L	1	08/28/2019 13:49
Method: SM 2540 D-1997									
Analyst: KMT									
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids									
Total Suspended Solids	ejj	A	1.1	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-04
Client Project:	Daily	Sampled:	08/27/2019 5:20
Client Sample ID:	001-Grab	Received:	08/28/2019 10:15
Sample Description:	001		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed	
			Method: EPA 1664B				Analyst: KMT			
Oil & Grease (HEM) by SPE										
Prep Date/Time: 08/28/2019 07:32										
Oil & Grease (HEM)	ejj	A	ND	1.4	5.0	U	mg/L	1	08/28/2019 14:12	

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-05
Client Project:	Daily	Sampled:	08/28/2019 5:37
Client Sample ID:	031-Grab	Received:	08/28/2019 10:15
Sample Description:	031		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: SM 9222 D-1997			Analyst: JAA			
Prep Date/Time: 08/28/2019 11:50									
Fecal Coliform by Membrane Filtration									
Fecal Coliform	d	A	4.0	1.0	1.0		CFU/100ml	1	08/28/2019 11:50
			Method: SM 5210 B-2001			Analyst: EF			
Prep Date/Time: 08/29/2019 15:27									
Biochemical Oxygen Demand									
Biochemical Oxygen Demand	ejj	A	2.6	2.0	2.0		mg/L	1	09/03/2019 22:07
			Method: SM 2540 D-1997			Analyst: KMT			
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids									
Total Suspended Solids	ejj	A	15	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-06
Client Project:	Daily	Sampled:	08/28/2019 5:40
Client Sample ID:	Mixed Liquor-Grab	Received:	08/28/2019 10:15
Sample Description:	Mixed Liquor		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed	
			Method: SM 2540 F-1997				Analyst: DAT			
										Prep Date/Time: 08/28/2019 10:56
Settleable Solids										
Settleable Solids	i	A	210	1.0	1.0		ml/L	1	08/28/2019 10:56	
			Method: SM 2540 D-1997				Analyst: KMT			
										Prep Date/Time: 08/28/2019 11:15
Total Suspended Solids										
Total Suspended Solids	ejj	A	2300	1.0	1.0		mg/L	1	08/28/2019 12:45	

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-07
Client Project:	Daily	Sampled:	08/28/2019 5:35
Client Sample ID:	J-Box-Grab	Received:	08/28/2019 10:15
Sample Description:	J-Box		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: EPA 350.1 Rev 2.0			Analyst: ABG			
Nitrogen, Ammonia as N			Prep Date/Time: 08/28/2019 11:20						
Nitrogen, Ammonia (As N)	ei	A	0.48	0.054	0.10		mg/L	1	08/28/2019 13:02
			Method: EPA 420.4 Rev 1.0			Analyst: ABG			
Total Phenolics			Prep Date/Time: 08/28/2019 11:10						
Phenolics, Total Recoverable	ejj	A	ND	0.0060	0.010	U	mg/L	1	08/28/2019 13:50
			Method: SM 2540 D-1997			Analyst: KMT			
Total Suspended Solids			Prep Date/Time: 08/28/2019 11:15						
Total Suspended Solids	ejj	A	10	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: *Wednesday, September 11, 2019*

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-08
Client Project:	Daily	Sampled:	08/28/2019 6:10
Client Sample ID:	WWII-Grab	Received:	08/28/2019 10:15
Sample Description:	WWII		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
Method: SM 4500-CN C/E-1999					Analyst: ABG				
Total Cyanide									
Prep Date/Time: 08/28/2019 11:45									
Cyanide, Total	ejj	A	0.019	0.0020	0.0050		mg/L	1	08/28/2019 14:42

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-09
Client Project:	Daily	Sampled:	08/28/2019 6:20
Client Sample ID:	Coldwell-Grab	Received:	08/28/2019 10:15
Sample Description:	Coldwell		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: EPA 200.7 Rev 4.4			Analyst: RPL			
Total Recoverable Metals by ICP						Prep Date/Time: 08/29/2019 08:23			
Lead	ejj	A	0.12	0.0033	0.0075		mg/L	1	08/30/2019 12:58
Zinc	ejj	A	0.55	0.0073	0.020		mg/L	1	08/30/2019 12:58
			Method: SM 4500-CN C/E-1999			Analyst: ABG			
Total Cyanide						Prep Date/Time: 08/28/2019 11:45			
Cyanide, Total	ejj	A	0.20	0.0020	0.0050		mg/L	1	08/28/2019 14:43
			Method: EPA 350.1 Rev 2.0			Analyst: ABG			
Nitrogen, Ammonia as N						Prep Date/Time: 08/28/2019 11:20			
Nitrogen, Ammonia (As N)	ei	A	59	0.54	1.0		mg/L	1	08/28/2019 13:04
			Method: SM 2540 D-1997			Analyst: KMT			
Total Suspended Solids						Prep Date/Time: 08/28/2019 11:15			
Total Suspended Solids	ejj	A	66	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-10
Client Project:	Daily	Sampled:	08/28/2019 6:25
Client Sample ID:	RSB FT Overflow-Grab	Received:	08/28/2019 10:15
Sample Description:	RSB FT Overflow		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: EPA 200.7 Rev 4.4			Analyst: RPL			
Total Recoverable Metals by ICP									
Prep Date/Time: 08/29/2019 08:23									
Lead	ejj	A	0.044	0.0033	0.0075		mg/L	1	08/30/2019 13:03
Zinc	ejj	A	0.091	0.0073	0.020		mg/L	1	08/30/2019 13:03
			Method: EPA 350.1 Rev 2.0			Analyst: ABG			
Nitrogen, Ammonia as N									
Prep Date/Time: 08/28/2019 11:20									
Nitrogen, Ammonia (As N)	ei	A	8.0	0.054	0.10		mg/L	1	08/28/2019 13:06
			Method: SM 2540 D-1997			Analyst: KMT			
Total Suspended Solids									
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids	ejj	A	14	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: *Wednesday, September 11, 2019*

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-11
Client Project:	Daily	Sampled:	08/28/2019 6:26
Client Sample ID:	RSB FT Influent-Grab	Received:	08/28/2019 10:15
Sample Description:	RSB FT Influent		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: SM 2540 D-1997			Analyst: KMT			
			Prep Date/Time: 08/28/2019 11:15						
Total Suspended Solids									
Total Suspended Solids	ejj	A	900	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: *Wednesday, September 11, 2019*

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-12
Client Project:	Daily	Sampled:	08/28/2019 6:48
Client Sample ID:	BFTD-Grab	Received:	08/28/2019 10:15
Sample Description:	BFTD		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed		
			Method: SM 2540 D-1997				Analyst: KMT				
										Prep Date/Time: 08/28/2019 11:15	
Total Suspended Solids											
Total Suspended Solids	ejj	A	52	1.0	1.0		mg/L	1	08/28/2019 12:45		

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-13
Client Project:	Daily	Sampled:	08/26/2019 7:58
Client Sample ID:	WPL-Grab	Received:	08/28/2019 10:15
Sample Description:	WPL		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed		
			Method: SM 4500 H+ B-2000				Analyst: DAT				
										Prep Date/Time: 08/29/2019 14:09	
pH											
pH	ejj	A	< 2		2.00	H	S.U.	1	08/29/2019 14:09		
			Method: SM 2710 F-2004				Analyst: DAT				
										Prep Date/Time: 08/28/2019 12:41	
Specific Gravity											
Specific Gravity		A	1.31	0.0100	0.0100		T/4 C	1	08/28/2019 12:41		

Analytical Results

Date: *Wednesday, September 11, 2019*

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-14
Client Project:	Daily	Sampled:	08/28/2019 6:32
Client Sample ID:	999-Grab	Received:	08/28/2019 10:15
Sample Description:	999		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: SM 2540 D-1997				Analyst: KMT		
Total Suspended Solids									
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids	ejj	A	2.2	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: *Wednesday, September 11, 2019*

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-15
Client Project:	Daily	Sampled:	08/28/2019 6:52
Client Sample ID:	BFTC-Grab	Received:	08/28/2019 10:15
Sample Description:	BFTC		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: SM 2540 D-1997			Analyst: KMT			
Total Suspended Solids									
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids	ejj	A	37	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-16
Client Project:	Daily	Sampled:	08/27/2019 6:56
Client Sample ID:	002-Composite	Received:	08/28/2019 10:15
Sample Description:	002		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: SM 4500-CN C/E-1999				Analyst: ABG		
Prep Date/Time: 08/28/2019 11:45									
Total Cyanide									
Cyanide, Total	ejj	A	ND	0.0020	0.0050	U	mg/L	1	08/28/2019 14:45
			Method: SM 2540 D-1997				Analyst: KMT		
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids									
Total Suspended Solids	ejj	A	1.6	1.0	1.0		mg/L	1	08/28/2019 12:45

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-17
Client Project:	Daily	Sampled:	08/27/2019 6:56
Client Sample ID:	002-Grab	Received:	08/28/2019 10:15
Sample Description:	002		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
Method: EPA 1664B					Analyst: KMT				
Oil & Grease (HEM) by SPE									
Prep Date/Time: 08/28/2019 07:32									
Oil & Grease (HEM)	ejj	A	ND	1.4	5.0	U	mg/L	1	08/28/2019 14:12

Analytical Results

Date: Wednesday, September 11, 2019

Client:	Arcelor Mittal USA, Inc.	Work Order/ID:	19H1769-18
Client Project:	Daily	Sampled:	08/27/2019 7:10
Client Sample ID:	WAL-Grab	Received:	08/28/2019 10:15
Sample Description:	WAL		
Matrix:	Aqueous		

Analyses	Certs	AT	Result	MDL	RL	Qual	Units	DF	Analyzed
			Method: EPA 1664B				Analyst: KMT		
Oil & Grease (HEM) by SPE									
Prep Date/Time: 08/28/2019 07:32									
Oil & Grease (HEM)	ejj	A	5.6	1.4	5.0		mg/L	1	08/28/2019 14:12
			Method: SM 2710 F-2004				Analyst: DAT		
Specific Gravity									
Prep Date/Time: 08/28/2019 12:41									
Specific Gravity		A	1.00	0.0100	0.0100		T/4 C	1	08/28/2019 12:41
			Method: SM 2540 D-1997				Analyst: KMT		
Total Suspended Solids									
Prep Date/Time: 08/28/2019 11:15									
Total Suspended Solids	ejj	A	25	1.0	1.0		mg/L	1	08/28/2019 12:45

ANALYTE TYPES: (AT)

A, B = Target Analyte

I = Internal Standard

M = Summation Analyte

S = Surrogate

T = Tentatively Identified Compound (TIC, concentration estimated)



QC SAMPLE IDENTIFICATIONS

BLK = Method Blank

DUP = Method Duplicate

BS = Method Blank Spike

MS = Matrix Spike

ICB = Initial Calibration Blank

CCB = Continuing Calibration Blank

CRL = Client Required Reporting Limit

PDS = Post Digestion Spike

QCS = Quality Control Standard

ICSA = Interference Check Standard "A"

ICSAB = Interference Check Standard "AB"

BSD = Method Blank Spike Duplicate

MSD = Matrix Spike Duplicate

ICV = Initial Calibration Verification

CCV = Continuing Calibration Verification

OPR = Ongoing Precision and Recovery Standard

SD = Serial Dilution

CERTIFICATIONS (Certs)

Below is a list of certifications maintained by the Microbac Merrillville Laboratory. All data included in this report has been reviewed for and meets all project specific and quality control requirements of the applicable accreditation, unless otherwise noted. Complete lists of individual analytes pursuant to each certification below are available upon request.

d Illinois EPA drinking water, wastewater and solid waste analysis (#200064)

e Illinois DOPH Micro analysis of drinking water (#1755266)

i Kansas Dept Health & Env. NELAP (#E-10397)

j Kentucky Wastewater Laboratory Certification Program (#108202)

FLAGS, FOOTNOTES AND ABBREVIATIONS (as needed)**H:** Sample was analyzed past holding time.**MDL:** Minimum Detection Limit**RL:** Reporting Limit**RPD:** Relative Percent Difference**U:** The analyte was analyzed for but was not detected above the reported quantitation limit. The quantitation limit has been adjusted for any dilution or concentration of the sample.

Cooler Receipt Log

Cooler ID: Default Cooler

Temp: 3.6°C
 MICROBAC®

Cooler Inspection Checklist

Ice Present or not required?	Yes
Shipping containers sealed or not required?	Yes
Custody seals intact or not required?	Yes
Chain of Custody (COC) Present?	Yes
COC includes customer information?	Yes
Relinquished and received signature on COC?	Yes
Sample collector identified on COC?	Yes
Sample type identified on COC?	Yes
Correct type of Containers Received	Yes
Correct number of containers listed on COC?	Yes
Containers Intact?	Yes
COC includes requested analyses?	Yes
Enough sample volume for indicated tests received?	Yes
Sample labels match COC (Name, Date & Time?)	Yes
Samples arrived within hold time?	Yes
Correct preservatives on COC or not required?	Yes
Chemical preservations checked or not required?	Yes
Preservation checks meet method requirements?	Yes
VOA vials have zero headspace, or not recd.?	Yes

Microbac Laboratories, Inc.

250 West 84th Drive | Merrillville, IN 46410 | 800.536.8379 p | 219.769.8378 p | 219.769.1664 f | www.microbac.com



Chain of Custody

ArcelorMittal Burns Harbor/Microbac Labs

Wednesday

Lab Work No: 19H1769

* Date Obtained 8-28-19

** Sample Date: 8-27-19

Location	Time	Sampler	Type	Preserved	Cooled	Containers			Parameters	Comments
						Type	Qty	Vol. (ml)		
011 **	<u>04:59</u>	<u>CP</u>	Comp	No	Yes	Glass	1	4000	NH3, TSS, Phenol, Zn, Cn, Pb	01
			Grab	No	No	Plastic	1	500	pH, Tot Res Cl	02
			Grab	Yes	No	Glass	1	1000	FOG (prepreserved)	↓
001 **	<u>05:20</u>		Comp	No	Yes	Glass	1	4000	NH3, Phenol, TSS	03
			Grab	No	Yes	Plastic	1	500	pH, Tot Res Cl	04
			Grab	Yes	No	Glass	1	1000	FOG (prepreserved)	↓
031 *	<u>05:37</u>		Grab	No	No	Plastic	1	1000	TSS	05
			Grab	No	No	Plastic	1	1000	BOD	↓
			Grab	Yes	No	Plastic	1	125	Fecal (sterilized bottle)	↓
Mixed Liquor *	<u>05:40</u>		Grab	No	No	Plastic	1	2000	TSS, Settling	06
J-Box *	<u>05:35</u>		Grab	No	No	Glass	2	1000	NH3, Phenol, TSS, pH	07
DIW-131 *	<u>NA</u>		Grab	No	No	Plastic	1	125	pH	X
WWII *	<u>06:10</u>		Grab	No	No	Plastic	1	1000	Cn	08
Coldwell	<u>06:20</u>		Grab	No	No	Plastic	2	2000	NH3, CN, Pb, Zn, TSS	09
RSB FT Overflow *	<u>06:25</u>		Grab	No	No	Plastic	2	1000	NH3, pH, TSS, Pb, Zn	10
RSB FT Influent *	<u>06:26</u>		Grab	No	No	Plastic	1	500	TSS	11
BFTD *	<u>06:48</u>		Grab	No	No	Plastic	1	500	TSS	12
WPL***	<u>07:58</u>		Grab	No	No	Glass	1	1000	SpG, pH	13
999 *	<u>06:32</u>		Grab	No	No	Plastic	1	500	TSS, pH	14
BFTC *	<u>06:52</u>		Grab	No	No	Plastic	1	500	TSS	15
002 **	<u>06:56</u>		Comp	No	Yes	Plastic	1	500	TSS	16
			Grab	No	No	Plastic	1	125	pH	17
			Grab	Yes	No	Glass	1	1000	FOG (prepreserved)	↓
WAL 1**	<u>07:10</u>		Grab	No	No	Glass	1	1000	TSS, SpG, pH	18
			Grab	Yes	No	Glass	2	1000	FOG (prepreserved)	↓ 19
WAL 2**	<u>5-D</u>		Grab	No	No	Glass	1	1000	TSS, SpG, pH	X
			Grab	Yes	No	Glass	2	1000	FOG (prepreserved)	X
WAL 3**	<u>07:10</u>		Grab	No	No	Glass	1	1000	TSS, SpG, pH	X
			Grab	Yes	No	Glass	2	1000	FOG (prepreserved)	X
SWTP *	<u>NA</u>	***	Grab	No	No	Plastic	10	1000	TSS	X

No GMS + HMs

*** WPL is for previous sample date

**** Sample collected by Water Process personnel

3.9

Relinquished by: C. Dulka

Date: 8-28-19

Time: 07:15

Received by: Hz. OTO

Date: 8/28/19

Time: 0800

Env 3x Rev. 15 04/27/17 (TEK)

19H1769 Carey Gadzala
ArcelorMittal - Burns Harbor, IN
Daily
08/28/2019



COMP IN FRIDGE

307258 Daily Work authorization form for all visiting workers

For each job, and before starting work at the job site, a contractor representative must meet face to face with the ArcelorMittal representative responsible for the work and discuss the work to be performed and any specific safety requirements.



The named contractor or work crew is cleared to perform the job described herein:

Section 1
 Company name Microbac Labs
 Company contact/phone no Cary Gatzela 769 8378
 location and project/job description Enviro Bldg / water Samples
 ArcelorMittal representative Wairin Hewitt
 ArcelorMittal representative department EB
 ArcelorMittal representative phone number 4863
 Date 8/28/14
 Call 4863
 Clinic pickup point 48

HIRAC-Lite	Yes	N/A	No	Yes	N/A	No	Yes	N/A	No
1) Are emergency evacuation areas identified and known?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	10) Could someone be caught in or between anything?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
2) Is there a current and valid isolation (LOTO) procedure?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	11) Could someone get hurt as a result of a fall from height?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
3) Will everyone apply a personal safety lock?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	12) Can something fall and/or strike me or someone else?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
4) Are there adjacent work crews exposed (including ArcelorMittal employees)?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	13) Is everyone properly trained for this job?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5) Are there potential hazards or high risk job steps?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	14) Are flags and deraills in place if needed?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
6) Do we have the correct tools for the job?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	15) Can we slip or trip on anything (including travel to and from the job)?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
7) Is additional PPE required?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	16) Have all affected people been notified?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
8) Is there a potential for exposure (chemical, radiation, laser, temperature)?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	17) Can we strain or overexert ourselves?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
9) Is someone working on or near energized electrical equipment (motor control rooms, overhead power lines, etc.)?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	18) Has equipment been inspected prior to use? (tools, PPE, mobile equipment, etc.)	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Other Hazards and Considerations for Discussion				Permits						
Yes	N/A	No		Yes	N/A	No	Yes	N/A	No	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	19) Pneumatic air tools & lines	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	37) Confined space	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	20) Vehicle / mob equip traffic	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	38) Energized electrical work	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	21) Gas hazards-CO, CO2, etc.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	39) Excavation / drilling	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	22) Hot process, metal, temp.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	40) Hot work	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	23) Pressurized / steam pipe	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	41) Other	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	24) Housekeeping	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	25) Production hazards	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	26) Material handling	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	27) Crane and rigging	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	28) Overhead work	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	29) Scaffold work	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	30) Explosives	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	31) Barricades	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	32) Radiation	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	33) Asbestos	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	34) Noise	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	35) Lasers	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	36) Sewers	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Section 3
 Visiting worker name (print) A. Otko Badge # 164042
 Hierarchy of Controls: 1. Elimination 2. Substitution 3. Engineering 4. Administrative 5. PPE
 Hazard # 15 Responsible Person B. Otko
 Controls Beams of uncut surfaces
17 Proper lifting of carts
20 Vehicle movement

My crew and I are familiar with the safety hazards/considerations for this job. We are prepared to perform the work in a safe "workmanship" like manner. I have reviewed these considerations with the ArcelorMittal representative named below.
 Contractor or crew leader [Signature] ArcelorMittal representative [Signature] Replacement rep/phone _____
 (Ensure form is fully completed prior to signing) Original to contractor, (1) copy to ArcelorMittal representative
 Controlled by Maintenance Administration Dept. ArcelorMittal Burns Harbor
 2016-04-BH-DailyWorkAuthorization

Microbac Laboratories - Chicagoland Division
pH - METHOD 9045D
Arcelor Mittal / Burns Harbor NPDES

Sample ID		pH	Analyst	Date/Time of Analysis
Buffer ID: Meter ID:	4: 185909	7: 188312	10: 187680	
Calibration	(4) (1) (1) (1)		DAO	8/27/19 0800
ICV	4 10	6.99		
Slope		100.7		
Lake 999		8.09 ^{9.97} 7.97		
Location 001		7.77		
Location 002		8.09		
Location 011		7.75		
WAL 1	_____	_____		
WAL 2	_____	_____		
SWTP J-Box		8.66		
DIW 131	_____	_____		
RSB		8.96		
Dup- 999		7.96		
CCV		7.01	↓	↓

Sample ID		pH	Analyst	Date/Time of Analysis
Buffer ID: Meter ID:	4: 185909	7: 188312	10: 187680	
Calibration	(4) (1) (1) (1)		DAO	8/28/19 0800
ICV	4 10	6.99		
Slope		101.3		
Lake 999		7.98		
Location 001		7.84		
Location 002		8.20		
Location 011		7.70		
WAL 1		8.93		
WAL 2	_____	_____		
SWTP J-Box		8.63		
DIW 131	_____	_____		
RSB		8.96		
Dup- 001		7.85		
CCV		7.01	↓	↓

Microbac Laboratories, Inc. - Chicagoland Division
Residual Chlorine - METHOD SM 4500-Cl I-2000
Arcelor Mittal /Burns Harbor NPDES

Meter ID: DH meter Residual Chlorine Standard: A 9074
 Iodine Reagent: _____ Acid Reagent: _____

Sample ID	Residual Chlorine	Analyst	Date/Time of Analysis
Cal-Std 1	0.02 mg/L	BAO	8/27/19 0800
Cal-Std 2	0.05 mg/L		
Cal-Std 3	0.1 mg/L		
Slope Blank	0.00		
LCS 0.02 mg/L	0.02		
011	0.00		
011 DUP	0.00		
001	0.00		
002	0.00		
003	0.00		
DUP 003	0.00		

Meter ID: BH meter Residual Chlorine Standard: A 9074
 Iodine Reagent: _____ Acid Reagent: _____

Sample ID	Residual Chlorine	Analyst	Date/Time of Analysis
Cal-Std 1	0.02 mg/L	BAO	8/28/19 0800
Cal-Std 2	0.05 mg/L		
Cal-Std 3	0.1 mg/L		
Slope Blank	0.00		
LCS 0.02 mg/L	0.10		
011	0.00		
011 DUP	0.00		
001	0.00		
002	0.00		
003	0.00		
DUP 001	0.00		

Meter ID: _____ Residual Chlorine Standard: _____
 Iodine Reagent: _____ Acid Reagent: _____

Sample ID	Residual Chlorine	Analyst	Date/Time of Analysis
Cal Std 1	0.02 mg/L		
Cal Std 2	0.05 mg/L		
Cal Std 3	0.1 mg/L		
Slope			
LCS 0.02 mg/L			
011			
011 DUP			
001			
002			
003			
DUP			